

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092525\
 Data File : VV039192.D
 Acq On : 25 Sep 2025 19:17
 Operator : SY/MD
 Sample : Q3195-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW-1-091725

Quant Time: Sep 26 01:37:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	535790	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	971101	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	956100	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	458217	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	427676	55.152	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	= 110.300%	
35) Dibromofluoromethane	4.503	113	398238	51.011	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	= 102.020%	
50) Toluene-d8	7.236	98	1062986	46.360	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	= 92.720%	
62) 4-Bromofluorobenzene	10.001	95	416581	44.133	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	= 88.260%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.298	62	2528	0.259	ug/l #	84
9) 1,1,2-Trichlorotrifluo...	2.082	101	6382	0.760	ug/l	97
12) 1,1-Dichloroethene	2.085	96	44339	5.489	ug/l	95
24) 1,1-Dichloroethane	3.117	63	6814	0.404	ug/l #	92
27) cis-1,2-Dichloroethene	3.825	96	154086	16.246	ug/l	88
32) 1,1,1-Trichloroethane	4.516	97	32296	2.116	ug/l	97
44) Trichloroethene	5.831	130	367152	41.396	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092525\
Data File : VV039192.D
Acq On : 25 Sep 2025 19:17
Operator : SY/MD
Sample : Q3195-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW-1-091725

Quant Time: Sep 26 01:37:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

