

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092525\
 Data File : VV039193.D
 Acq On : 25 Sep 2025 19:40
 Operator : SY/MD
 Sample : Q3195-09
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW-22-091825

Quant Time: Sep 26 01:38:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.603	168	525270	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	956838	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	943207	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	451495	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	428579	56.376	ug/l	0.00
Spiked Amount 50.000	Range 81 - 118		Recovery = 112.760%			
35) Dibromofluoromethane	4.503	113	392109	50.975	ug/l	0.00
Spiked Amount 50.000	Range 80 - 119		Recovery = 101.940%			
50) Toluene-d8	7.239	98	1050090	46.480	ug/l	0.00
Spiked Amount 50.000	Range 89 - 112		Recovery = 92.960%			
62) 4-Bromofluorobenzene	10.001	95	412268	44.327	ug/l	0.00
Spiked Amount 50.000	Range 85 - 114		Recovery = 88.660%			
Target Compounds						
30) Chloroform	4.291	83	16860	0.975	ug/l	100
38) Carbon Tetrachloride	4.738	117	24918m	1.753	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092525\
Data File : VV039193.D
Acq On : 25 Sep 2025 19:40
Operator : SY/MD
Sample : Q3195-09
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW-22-091825

Quant Time: Sep 26 01:38:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

