

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092525\
 Data File : VV039195.D
 Acq On : 25 Sep 2025 20:25
 Operator : SY/MD
 Sample : Q3195-11
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW-24-091925

Quant Time: Sep 26 01:39:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	515327	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	928629	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	925762	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	442595	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	414441	55.568	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	111.140%
35) Dibromofluoromethane	4.503	113	384300	51.477	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	102.960%
50) Toluene-d8	7.236	98	1012096	46.160	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	92.320%
62) 4-Bromofluorobenzene	10.002	95	401201	44.448	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	88.900%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.121	85	9372	1.149	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.089	101	8698	1.077	ug/l	99
30) Chloroform	4.298	83	5926	0.349	ug/l #	83
44) Trichloroethene	5.851	130	8633	1.018	ug/l	92
64) Tetrachloroethene	7.902	164	2650	0.342	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092525\
Data File : VV039195.D
Acq On : 25 Sep 2025 20:25
Operator : SY/MD
Sample : Q3195-11
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW-24-091925

Quant Time: Sep 26 01:39:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

