

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
 Data File : VV039199.D
 Acq On : 26 Sep 2025 10:00
 Operator : SY/MD
 Sample : VV0926MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV0926MBL01

Quant Time: Sep 27 02:49:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.596	168	545248	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	960075	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.773	117	938706	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	463545	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.937	65	423576	53.676	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	107.360%
35) Dibromofluoromethane	4.497	113	390719	50.623	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	101.240%
50) Toluene-d8	7.236	98	1049117	46.281	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	92.560%
62) 4-Bromofluorobenzene	10.001	95	406277	43.536	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	87.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
Data File : VV039199.D
Acq On : 26 Sep 2025 10:00
Operator : SY/MD
Sample : VV0926MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV0926MBL01

Quant Time: Sep 27 02:49:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

