

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
 Data File : VV039200.D
 Acq On : 26 Sep 2025 10:23
 Operator : SY/MD
 Sample : VV0926WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV0926WBL01

Quant Time: Sep 27 02:50:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	488972	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	870889	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	870759	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	422120	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	399373	56.433	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	= 112.860%	
35) Dibromofluoromethane	4.503	113	357344	51.040	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	= 102.080%	
50) Toluene-d8	7.236	98	952055	46.300	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	= 92.600%	
62) 4-Bromofluorobenzene	10.001	95	387045	45.722	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	= 91.440%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
Data File : VV039200.D
Acq On : 26 Sep 2025 10:23
Operator : SY/MD
Sample : VV0926WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV0926WBL01

Quant Time: Sep 27 02:50:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

