

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
 Data File : VV039204.D
 Acq On : 26 Sep 2025 11:58
 Operator : SY/MD
 Sample : Q3195-12
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW-26-091925

Quant Time: Sep 27 02:53:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.603	168	486629	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	867427	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.777	117	873266	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	420754	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	394452	56.006	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	112.020%
35) Dibromofluoromethane	4.503	113	367761	52.738	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	105.480%
50) Toluene-d8	7.240	98	962872	47.013	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	94.020%
62) 4-Bromofluorobenzene	10.002	95	378432	44.883	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	89.760%
Target Compounds						
30) Chloroform	4.288	83	18760	1.172	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
Data File : VV039204.D
Acq On : 26 Sep 2025 11:58
Operator : SY/MD
Sample : Q3195-12
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW-26-091925

Quant Time: Sep 27 02:53:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

