

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
 Data File : VV039205.D
 Acq On : 26 Sep 2025 12:21
 Operator : SY/MD
 Sample : Q3195-04 100X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW-4-091625

Quant Time: Sep 27 02:53:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	513652	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	924164	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	911885	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	438479	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.940	65	416085	55.970	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	111.940%
35) Dibromofluoromethane	4.503	113	392713	52.858	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	105.720%
50) Toluene-d8	7.239	98	1008875	46.235	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	92.460%
62) 4-Bromofluorobenzene	10.001	95	396215	44.107	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	88.220%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.082	96	11345	1.465	ug/l	96
27) cis-1,2-Dichloroethene	3.825	96	144878	15.933	ug/l	87
44) Trichloroethene	5.831	130	804569	95.322	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
Data File : VV039205.D
Acq On : 26 Sep 2025 12:21
Operator : SY/MD
Sample : Q3195-04 100X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW-4-091625

Quant Time: Sep 27 02:53:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

