

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
 Data File : VV039209.D
 Acq On : 26 Sep 2025 13:51
 Operator : SY/MD
 Sample : Q3195-06
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW-19D-092225

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025
 Supervised By :Semsettin Yesilyurt 09/29/2025

Quant Time: Sep 27 02:55:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.603	168	344932	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	614959	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	627934	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	291769	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	308774	61.851	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	= 123.700%#	
35) Dibromofluoromethane	4.503	113	284583	57.564	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	= 115.120%	
50) Toluene-d8	7.240	98	686292	47.266	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	= 94.540%	
62) 4-Bromofluorobenzene	10.002	95	261604	43.765	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	= 87.540%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.121	85	32529	5.957	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.082	101	41397	7.658	ug/l	100
27) cis-1,2-Dichloroethene	3.847	96	3033m	0.497	ug/l	
30) Chloroform	4.291	83	7336	0.646	ug/l	98
44) Trichloroethene	5.847	130	18064	3.216	ug/l	93
64) Tetrachloroethene	7.908	164	1575	0.299	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
Data File : VV039209.D
Acq On : 26 Sep 2025 13:51
Operator : SY/MD
Sample : Q3195-06
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW-19D-092225

Quant Time: Sep 27 02:55:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 09/29/2025
Supervised By :Semsettin Yesilyurt 09/29/2025

