

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
 Data File : VV039210.D
 Acq On : 26 Sep 2025 14:13
 Operator : SY/MD
 Sample : Q3195-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW-19M-092325

Quant Time: Sep 27 02:56:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.603	168	348105	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	639098	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	643573	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	305861	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	317196	62.959	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	125.920%#
35) Dibromofluoromethane	4.503	113	295601	57.534	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	115.060%
50) Toluene-d8	7.239	98	700315	46.410	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	92.820%
62) 4-Bromofluorobenzene	10.001	95	266839	42.955	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	85.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
Data File : VV039210.D
Acq On : 26 Sep 2025 14:13
Operator : SY/MD
Sample : Q3195-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW-19M-092325

Quant Time: Sep 27 02:56:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

