

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
 Data File : VV039211.D
 Acq On : 26 Sep 2025 14:36
 Operator : SY/MD
 Sample : Q3195-08
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW-21M-091925

Quant Time: Sep 27 02:56:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	513487	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	944501	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	957014	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	451121	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	452207	60.849	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	121.700%#
35) Dibromofluoromethane	4.500	113	418436	55.108	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	110.220%
50) Toluene-d8	7.236	98	1055705	47.339	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	94.680%
62) 4-Bromofluorobenzene	10.001	95	402477	43.840	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	87.680%
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	3.834	96	16420	1.806	ug/l	82
44) Trichloroethene	5.844	130	17696	2.051	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
Data File : VV039211.D
Acq On : 26 Sep 2025 14:36
Operator : SY/MD
Sample : Q3195-08
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW-21M-091925

Quant Time: Sep 27 02:56:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

