

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
 Data File : VV039212.D
 Acq On : 26 Sep 2025 14:58
 Operator : SY/MD
 Sample : Q3195-13
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW-26-091925-FD

Quant Time: Sep 27 02:57:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	454191	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	813227	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	842320	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	398907	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	398953	60.691	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	121.380%#
35) Dibromofluoromethane	4.503	113	366722	56.094	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	112.180%
50) Toluene-d8	7.239	98	915207	47.664	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	95.320%
62) 4-Bromofluorobenzene	10.001	95	350373	44.325	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	88.660%
Target Compounds						
30) Chloroform	4.288	83	17020	1.139	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
Data File : VV039212.D
Acq On : 26 Sep 2025 14:58
Operator : SY/MD
Sample : Q3195-13
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW-26-091925-FD

Quant Time: Sep 27 02:57:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

