

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
 Data File : VV039213.D
 Acq On : 26 Sep 2025 15:20
 Operator : SY/MD
 Sample : Q3195-14
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW-29-092325

Quant Time: Sep 27 02:58:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	450640	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	819940	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	841715	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	391674	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.941	65	402866	61.769	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	123.540%#
35) Dibromofluoromethane	4.503	113	368992	55.979	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	111.960%
50) Toluene-d8	7.239	98	896272	46.296	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	92.600%
62) 4-Bromofluorobenzene	10.001	95	348361	43.710	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	87.420%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.089	101	2185	0.309	ug/l	# 85
12) 1,1-Dichloroethene	2.082	96	2472	0.364	ug/l	83
27) cis-1,2-Dichloroethene	3.834	96	13996	1.754	ug/l	81
44) Trichloroethene	5.844	130	21362	2.853	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
Data File : VV039213.D
Acq On : 26 Sep 2025 15:20
Operator : SY/MD
Sample : Q3195-14
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW-29-092325

Quant Time: Sep 27 02:58:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

