

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092625\  
 Data File : VV039217.D  
 Acq On : 26 Sep 2025 16:50  
 Operator : SY/MD  
 Sample : Q3195-18  
 Misc : 5.0mL/MSVOA\_V/WATER  
 ALS Vial : 21 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 MW-32S-091725

Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 10/28/2025  
 Supervised By :mohammad ahmed 10/28/2025

Quant Time: Sep 29 07:53:34 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 24 08:08:08 2025  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response             | Conc   | Units  | Dev(Min) |
|------------------------------|----------------|------|----------------------|--------|--------|----------|
| -----                        |                |      |                      |        |        |          |
| Internal Standards           |                |      |                      |        |        |          |
| 1) Pentafluorobenzene        | 4.603          | 168  | 339546               | 50.000 | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 5.535          | 114  | 617241               | 50.000 | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 8.777          | 117  | 641272               | 50.000 | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 11.175         | 152  | 293243               | 50.000 | ug/l   | 0.00     |
| System Monitoring Compounds  |                |      |                      |        |        |          |
| 33) 1,2-Dichloroethane-d4    | 4.944          | 65   | 306687               | 62.408 | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 81 - 118 |      | Recovery = 124.820%# |        |        |          |
| 35) Dibromofluoromethane     | 4.503          | 113  | 290412               | 58.526 | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 80 - 119 |      | Recovery = 117.060%  |        |        |          |
| 50) Toluene-d8               | 7.240          | 98   | 683404               | 46.893 | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 89 - 112 |      | Recovery = 93.780%   |        |        |          |
| 62) 4-Bromofluorobenzene     | 10.002         | 95   | 255689               | 42.617 | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 85 - 114 |      | Recovery = 85.240%   |        |        |          |
| Target Compounds             |                |      |                      |        |        |          |
|                              |                |      |                      |        | Qvalue |          |
| 9) 1,1,2-Trichlorotrifluo... | 2.085          | 101  | 33493                | 6.294  | ug/l   | 99       |
| 27) cis-1,2-Dichloroethene   | 3.831          | 96   | 8375m                | 1.393  | ug/l   |          |
| 44) Trichloroethene          | 5.854          | 130  | 9552                 | 1.694  | ug/l   | 88       |
| 64) Tetrachloroethene        | 7.905          | 164  | 6243                 | 1.162  | ug/l   | 89       |
| -----                        |                |      |                      |        |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV092625\  
Data File : VV039217.D  
Acq On : 26 Sep 2025 16:50  
Operator : SY/MD  
Sample : Q3195-18  
Misc : 5.0mL/MSVOA\_V/WATER  
ALS Vial : 21 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
MW-32S-091725

Quant Time: Sep 29 07:53:34 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\82V092225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 24 08:08:08 2025  
Response via : Initial Calibration

Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 10/28/2025  
Supervised By :mohammad ahmed 10/28/2025

