

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
 Data File : VV039218.D
 Acq On : 26 Sep 2025 17:13
 Operator : SY/MD
 Sample : Q3195-19
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW-33D-091725

Quant Time: Sep 29 07:54:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.603	168	365701	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	671850	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	688047	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	318013	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	336265	63.533	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	= 127.060%#	
35) Dibromofluoromethane	4.503	113	314137	58.161	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	= 116.320%	
50) Toluene-d8	7.239	98	747340	47.112	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	= 94.220%	
62) 4-Bromofluorobenzene	10.001	95	282200	43.213	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	= 86.420%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.121	85	60345	10.424	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.085	101	30049	5.243	ug/l	99
12) 1,1-Dichloroethene	2.085	96	2302	0.417	ug/l #	71
27) cis-1,2-Dichloroethene	3.838	96	10444	1.613	ug/l	45
44) Trichloroethene	5.841	130	41354	6.739	ug/l	96
64) Tetrachloroethene	7.902	164	9610	1.668	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092625\
Data File : VV039218.D
Acq On : 26 Sep 2025 17:13
Operator : SY/MD
Sample : Q3195-19
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW-33D-091725

Quant Time: Sep 29 07:54:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

