

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092718\
 Data File : VV007916.D
 Acq On : 27 Sep 2018 19:23
 Operator : SY/MD
 Sample : J5063-19MS
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GAC39MS

Quant Time: Sep 28 03:39:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 28 00:42:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	87255	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	83033	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	40664	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	22862	4.89	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.80%
7) Chloroethane-d5	1.59	69	21469	5.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.20%
11) 1,1-Dichloroethene-d2	2.14	63	53103	5.03	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.60%
20) 2-Butanone-d5	3.96	46	67852	48.23	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.46%
24) Chloroform-d	4.41	84	53644	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.60%
26) 1,2-Dichloroethane-d4	5.09	65	29335	5.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.20%
32) Benzene-d6	5.10	84	101857	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.80%
36) 1,2-Dichloropropane-d6	6.12	67	32536	5.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.60%
41) Toluene-d8	7.36	98	96495	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.60%
43) trans-1,3-Dichloropropene-	7.67	79	10615	4.39	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.80%
46) 2-Hexanone-d5	8.14	63	50729	47.50	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.00%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	23749	4.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	40443	5.14	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.80%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
12) 1,1-Dichloroethene	2.14	96	29041	5.035	ug/L	96
25) Chloroform	4.43	83	12817	1.092	ug/L	98
31) Carbon tetrachloride	4.88	117	20412	2.319	ug/L	98
33) Benzene	5.15	78	122404	5.148	ug/L	100
34) Trichloroethene	5.96	95	34265	5.033	ug/L	99
42) Toluene	7.43	91	133109	5.141	ug/L	99
51) Chlorobenzene	8.93	112	88184	4.931	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092718\
Data File : VV007916.D
Acq On : 27 Sep 2018 19:23
Operator : SY/MD
Sample : J5063-19MS
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAC39MS

Quant Time: Sep 28 03:39:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 28 00:42:21 2018
Response via : Initial Calibration

