

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV092718\
 Data File : VV007933.D
 Acq On : 28 Sep 2018 03:08
 Operator : SY/MD
 Sample : J5062-03MSD
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GAC03MSD

Quant Time: Oct 11 12:17:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 28 03:46:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	88440	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	83576	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	40707	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	20666	4.36	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.20%
7) Chloroethane-d5	1.59	69	20401	4.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.60%
11) 1,1-Dichloroethene-d2	2.14	63	50237	4.69	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.80%
20) 2-Butanone-d5	3.96	46	67512	47.34	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.68%
24) Chloroform-d	4.41	84	51280	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.00%
26) 1,2-Dichloroethane-d4	5.09	65	28568	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
32) Benzene-d6	5.10	84	98213	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.40%
36) 1,2-Dichloropropane-d6	6.12	67	31353	4.96	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.20%
41) Toluene-d8	7.36	98	91948	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.60%
43) trans-1,3-Dichloropropene-	7.67	79	9872	4.05	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.00%
46) 2-Hexanone-d5	8.14	63	47571	44.25	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.50%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	22602	4.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	38828	4.93	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.60%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
12) 1,1-Dichloroethene	2.14	96	29167	4.989	ug/L	96
13) Acetone	2.21	43	4785	4.605	ug/L	95
31) Carbon tetrachloride	4.88	117	114253	12.896	ug/L	99
33) Benzene	5.15	78	121867	5.092	ug/L	100
34) Trichloroethene	5.96	95	32573	4.753	ug/L	98
42) Toluene	7.44	91	130970	5.025	ug/L	98
51) Chlorobenzene	8.93	112	86804	4.822	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV092718\
Data File : VV007933.D
Acq On : 28 Sep 2018 03:08
Operator : SY/MD
Sample : J5062-03MSD
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAC03MSD

Quant Time: Oct 11 12:17:41 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 28 03:46:50 2018
Response via : Initial Calibration

