


Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092718\
Data File : VV007896.D
Acq On : 27 Sep 2018 09:02
Operator : SY/MD
Sample : BFB76
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB76

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
Title : TRACE VOA SOM01.0
Last Update : Fri Sep 28 03:46:50 2018



AutoFind: Scans 2808, 2809, 2810; Background Corrected with Scan 2796

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	7327	PASS
75	95	30	80	51.7	20042	PASS
95	95	100	100	100.0	38786	PASS
96	95	5	9	6.8	2656	PASS
173	174	0.00	2	0.5	193	PASS
174	95	50	120	96.7	37504	PASS
175	174	5	9	8.3	3126	PASS
176	174	95	101	95.7	35904	PASS
177	176	5	9	6.7	2423	PASS

Scan 2796 (10.080 min): VV007896.D
BFB76

m/z	Abundance
39.90	144.0
43.90	233.0

Instrument :
MSVOA_V
ClientSampleId :
BFB76

m/z	Abundance
36.00	350.0
37.00	1971.0
38.00	1584.0
38.90	634.0
39.90	115.0
43.80	519.0
44.90	411.0
47.00	421.0
47.80	256.0
49.00	1371.0
50.00	6954.0
51.00	2180.0
52.00	103.0
54.90	139.0
55.90	626.0
56.90	1056.0
60.00	287.0
61.00	1853.0
62.00	1836.0
63.00	1456.0
63.90	210.0
64.90	324.0
66.70	105.0
68.00	3909.0
69.00	3996.0
69.90	282.0
71.80	222.0
73.00	1780.0
74.00	6335.0
75.00	18912.0
75.90	1675.0
76.90	164.0
77.80	149.0
78.10	149.0
78.90	1214.0
79.90	470.0
80.80	1323.0
81.80	346.0
86.90	1348.0
87.90	1238.0
90.90	183.0
92.00	1157.0
93.00	1769.0
94.00	4548.0
95.00	36656.0
96.00	2635.0
103.90	180.0
105.90	198.0
115.90	134.0
116.80	330.0
117.80	145.0
118.80	236.0
127.80	165.0
129.80	177.0
140.80	472.0
142.90	527.0
171.80	386.0
173.90	34384.0
174.90	2864.0
175.90	32576.0
176.90	2281.0

Instrument :
MSVOA_V
ClientSampleId :
BFB76

m/z	Abundance
36.00	365.0
37.00	1980.0
38.00	1709.0
39.00	687.0
39.90	120.0
43.80	513.0
44.90	447.0
47.00	409.0
47.80	291.0
49.00	1582.0
50.00	7595.0
51.00	2345.0
54.80	146.0
55.90	697.0
56.90	1195.0
60.00	309.0
61.00	2034.0
62.00	2005.0
63.00	1608.0
63.90	238.0
64.90	348.0
66.90	126.0
68.00	4274.0
69.00	4364.0
69.90	342.0
71.90	271.0
73.00	1948.0
74.00	7039.0
75.00	20736.0
76.00	1849.0
76.90	191.0
77.80	188.0
78.80	1329.0
79.90	517.0
80.90	1509.0
81.80	400.0
86.90	1593.0
87.90	1370.0
90.90	210.0
92.00	1240.0
93.00	1959.0
94.00	5030.0
95.00	40008.0
96.00	2728.0
103.90	174.0
105.90	223.0
115.80	169.0
116.80	341.0
117.90	139.0
118.90	241.0
127.80	176.0
129.80	194.0
140.80	549.0
142.90	560.0
145.80	105.0
147.80	106.0
154.80	105.0
171.80	487.0
173.90	38640.0
175.00	3256.0
175.90	36960.0
176.90	2508.0

Instrument :
MSVOA_V
ClientSampleId :
BFB76

m/z	Abundance
36.00	349.0
37.00	1920.0
38.00	1659.0
39.00	726.0
39.80	124.0
43.80	451.0
44.90	409.0
47.00	385.0
47.80	289.0
49.00	1621.0
50.00	7432.0
51.00	2223.0
54.90	155.0
55.90	745.0
57.00	1187.0
60.00	298.0
61.00	1991.0
62.00	2036.0
63.00	1597.0
63.90	197.0
64.90	337.0
67.00	144.0
68.00	4157.0
68.90	4309.0
70.00	368.0
71.90	272.0
73.00	1936.0
74.00	7030.0
75.00	20480.0
76.00	1760.0
77.00	217.0
77.80	210.0
78.80	1344.0
79.90	492.0
80.90	1599.0
81.80	401.0
86.90	1646.0
87.90	1331.0
90.90	215.0
91.90	1231.0
93.00	1947.0
94.00	5218.0
95.00	39696.0
96.00	2606.0
103.90	167.0
104.90	101.0
105.90	212.0
115.80	181.0
116.80	327.0
117.90	134.0
118.90	236.0
127.90	176.0
128.80	104.0
129.80	181.0
140.80	568.0
142.90	533.0
145.80	108.0
147.70	107.0
154.90	101.0
171.80	544.0
173.00	580.0
173.90	39488.0
175.00	3260.0
175.90	38176.0
176.90	2481.0

Instrument :
MSVOA_V
ClientSampleId :
BFB76