

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092724\
 Data File : VV037616.D
 Acq On : 27 Sep 2024 19:13
 Operator : SY/MD
 Sample : VV0927MBL02
 Misc : 5.0G/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK218

Quant Time: Sep 28 00:49:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 28 00:49:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	636517	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	609541	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	288069	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	215249	38.523	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.040%
7) Chloroethane-d5	1.532	69	185854	42.852	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.700%
11) 1,1-Dichloroethene-d2	2.060	65	99728	42.720	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.440%
21) 2-Butanone-d5	3.796	46	211351	106.196	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.200%
24) Chloroform-d	4.243	84	475790	47.861	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.720%
26) 1,2-Dichloroethane-d4	4.937	65	294248	50.569	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.140%
32) Benzene-d6	4.953	84	921727	48.324	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.640%
36) 1,2-Dichloropropane-d6	5.982	67	301506	49.642	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.280%
41) Toluene-d8	7.236	98	822846	47.576	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.160%
43) trans-1,3-Dichloroprop...	7.551	79	135751	48.114	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.220%
47) 2-Hexanone-d5	8.024	63	132160	103.048	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.050%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	361105	51.169	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.340%
66) 1,2-Dichlorobenzene-d4	11.554	152	326991	53.868	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.740%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092724\
Data File : VV037616.D
Acq On : 27 Sep 2024 19:13
Operator : SY/MD
Sample : VV0927MBL02
Misc : 5.0G/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK218

Quant Time: Sep 28 00:49:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 28 00:49:19 2024
Response via : Initial Calibration

