

Data File : VV007976.D
Acq On : 29 Sep 2018 01:51
Operator : SY/MD
Sample : J5091-07
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E62T1

Quant Time: Oct 10 04:10:48 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Oct 01 04:55:39 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	88611	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	84012	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	42312	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	17809	3.75	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.00%
7) Chloroethane-d5	1.59	69	18057	4.40	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.00%
11) 1,1-Dichloroethene-d2	2.14	63	31187	2.91	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.20%#
20) 2-Butanone-d5	3.97	46	69599	48.71	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.42%
24) Chloroform-d	4.41	84	72594	6.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	134.20%#
26) 1,2-Dichloroethane-d4	5.09	65	27109	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.20%
32) Benzene-d6	5.10	84	92920	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.60%
36) 1,2-Dichloropropane-d6	6.12	67	30769	4.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.80%
41) Toluene-d8	7.36	98	84436	4.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.60%
43) trans-1,3-Dichloropropene-	7.67	79	9470	3.87	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.40%
46) 2-Hexanone-d5	8.14	63	49784	46.07	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.14%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	23579	4.81	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	38247	4.67	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.40%

Target Compounds

						Qvalue
25) Chloroform	4.43	83	555503	46.623	ug/L	99
31) Carbon tetrachloride	4.88	117	447821	50.284	ug/L	100
51) Chlorobenzene	8.93	112	63495	3.509	ug/L	100
62) 1,3-Dichlorobenzene	11.23	146	36464	2.610	ug/L	97
63) 1,4-Dichlorobenzene	11.32	146	62811	4.325	ug/L	98
65) 1,2-Dichlorobenzene	11.69	146	76979	5.846	ug/L	98
68) 1,2,4-trichlorobenzene	13.32	180	10141	1.003	ug/L	97
70) 1,2,3-Trichlorobenzene	13.80	180	3076	0.336	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV007976.D
Acq On : 29 Sep 2018 01:51
Operator : SY/MD
Sample : J5091-07
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E62T1

Quant Time: Oct 10 04:10:48 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Oct 01 04:55:39 2018
Response via : Initial Calibration

