

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092820\
 Data File : VV018489.D
 Acq On : 28 Sep 2020 13:26
 Operator : SY/MD
 Sample : L4108-08
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Sep 29 02:44:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Sep 29 00:50:40 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	354130	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	337864	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	170098	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	124752	43.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.46%
7) Chloroethane-d5	1.58	69	105646	44.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.48%
11) 1,1-Dichloroethene-d2	2.12	63	176935	33.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.84%
21) 2-Butanone-d5	3.92	46	172788	97.82	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.82%
24) Chloroform-d	4.38	84	224615	44.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.72%
26) 1,2-Dichloroethane-d4	5.06	65	148057	46.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.26%
32) Benzene-d6	5.08	84	465643	48.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.28%
36) 1,2-Dichloropropane-d6	6.09	67	148841	48.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.86%
41) Toluene-d8	7.34	98	422339	47.91	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.82%
43) trans-1,3-Dichloropropene-	7.64	79	73047	46.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.54%
47) 2-Hexanone-d5	8.11	63	107324	92.19	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.19%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	179770	46.61	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.22%
64) 1,2-Dichlorobenzene-d4	11.65	152	176862	50.63	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.26%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092820\
Data File : VV018489.D
Acq On : 28 Sep 2020 13:26
Operator : SY/MD
Sample : L4108-08
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Sep 29 02:44:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 29 00:50:40 2020
Response via : Initial Calibration

