

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092821\
 Data File : VW022366.D
 Acq On : 28 Sep 2021 14:31
 Operator : SY/MD
 Sample : M3876-05
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW7Z1

Quant Time: Sep 29 01:34:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 29 01:30:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	266961	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	253209	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	111372	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	98515	43.483	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	86.960%		
7) Chloroethane-d5	1.561	69	82173	45.857	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	91.720%		
11) 1,1-Dichloroethene-d2	2.105	63	134066	34.035	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	68.080%		
21) 2-Butanone-d5	3.873	46	141501	88.499	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	88.500%		
24) Chloroform-d	4.349	84	179996	43.998	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	88.000%		
26) 1,2-Dichloroethane-d4	5.031	65	118318	47.162	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.320%		
32) Benzene-d6	5.050	84	390518	49.025	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	98.040%		
36) 1,2-Dichloropropane-d6	6.069	67	126780	49.749	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	99.500%		
41) Toluene-d8	7.317	98	346447	47.517	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	95.040%		
43) trans-1,3-Dichloroprop...	7.622	79	53630	46.087	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	92.180%		
47) 2-Hexanone-d5	8.088	63	93419	90.216	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	90.220%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	174032	47.045	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	94.100%		
66) 1,2-Dichlorobenzene-d4	11.625	152	133946	54.087	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	108.180%		
Target Compounds						Qvalue
13) Acetone	2.163	43	7495	5.973	ug/L	98
25) Chloroform	4.381	83	12542	3.331	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092821\
Data File : VV022366.D
Acq On : 28 Sep 2021 14:31
Operator : SY/MD
Sample : M3876-05
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW7Z1

Quant Time: Sep 29 01:34:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092721WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 29 01:30:11 2021
Response via : Initial Calibration

