

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092920\
 Data File : VV018533.D
 Acq On : 29 Sep 2020 22:13
 Operator : SY/MD
 Sample : L4149-14
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3N8

Quant Time: Sep 30 05:09:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 30 04:50:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	408751	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	394042	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	200141	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	111650	46.57	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.14%
7) Chloroethane-d5	1.58	69	100989	50.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.90%
11) 1,1-Dichloroethene-d2	2.12	63	178555	36.56	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.12%
21) 2-Butanone-d5	3.91	46	181705	101.62	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.62%
24) Chloroform-d	4.38	84	238041	45.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.84%
26) 1,2-Dichloroethane-d4	5.06	65	158817	48.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.52%
32) Benzene-d6	5.08	84	492990	49.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.78%
36) 1,2-Dichloropropane-d6	6.09	67	155420	51.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.08%
41) Toluene-d8	7.34	98	455774	48.54	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.08%
43) trans-1,3-Dichloropropene-	7.64	79	76437	46.92	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.84%
47) 2-Hexanone-d5	8.11	63	118068	103.07	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.07%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	193831	48.48	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.96%
64) 1,2-Dichlorobenzene-d4	11.65	152	199070	51.72	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.44%

Target Compounds

					Ovalue
17) trans-1,2-Dichloroethene	2.78	96	2540	0.870 ug/L	93
20) cis-1,2-Dichloroethene	3.94	96	7438	2.397 ug/L	87
34) Trichloroethene	5.94	95	1281220	381.222 ug/L	99
46) Tetrachloroethene	8.00	164	3987048	1462.295 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092920\
Data File : VV018533.D
Acq On : 29 Sep 2020 22:13
Operator : SY/MD
Sample : L4149-14
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3N8

Quant Time: Sep 30 05:09:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 30 04:50:47 2020
Response via : Initial Calibration

