

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092920\
 Data File : VV018546.D
 Acq On : 30 Sep 2020 03:20
 Operator : SY/MD
 Sample : L4165-04
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AA4

Quant Time: Sep 30 05:55:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 30 04:50:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	367588	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	358584	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	181450	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	94973	44.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.10%
7) Chloroethane-d5	1.58	69	91636	51.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	102.82%
11) 1,1-Dichloroethene-d2	2.12	63	153988	35.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.12%
21) 2-Butanone-d5	3.92	46	160041	99.52	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.52%
24) Chloroform-d	4.38	84	210388	45.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.26%
26) 1,2-Dichloroethane-d4	5.06	65	141911	47.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.90%
32) Benzene-d6	5.08	84	430607	47.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.82%
36) 1,2-Dichloropropane-d6	6.09	67	133710	48.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.44%
41) Toluene-d8	7.34	98	393789	46.08	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.16%
43) trans-1,3-Dichloropropene-	7.64	79	64811	43.72	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.44%
47) 2-Hexanone-d5	8.11	63	89389	85.75	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.75%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	160548	44.13	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.26%
64) 1,2-Dichlorobenzene-d4	11.65	152	174490	50.00	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092920\
Data File : VV018546.D
Acq On : 30 Sep 2020 03:20
Operator : SY/MD
Sample : L4165-04
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AA4

Quant Time: Sep 30 05:55:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 30 04:50:47 2020
Response via : Initial Calibration

