

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092920\
 Data File : VV018550.D
 Acq On : 30 Sep 2020 04:54
 Operator : SY/MD
 Sample : L4165-08
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AA8

Quant Time: Sep 30 06:09:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 30 04:50:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	352842	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	345496	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	174103	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	88986	43.00	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.00%
7) Chloroethane-d5	1.58	69	86645	50.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.28%
11) 1,1-Dichloroethene-d2	2.12	63	151501	35.93	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.86%
21) 2-Butanone-d5	3.92	46	150908	97.77	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.77%
24) Chloroform-d	4.38	84	203587	45.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.00%
26) 1,2-Dichloroethane-d4	5.06	65	136454	48.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.08%
32) Benzene-d6	5.07	84	409786	46.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.66%
36) 1,2-Dichloropropane-d6	6.10	67	131611	49.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.54%
41) Toluene-d8	7.34	98	377680	45.87	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.74%
43) trans-1,3-Dichloropropene-	7.64	79	62276	43.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.20%
47) 2-Hexanone-d5	8.11	63	96353	95.94	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.94%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	167808	47.87	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.74%
64) 1,2-Dichlorobenzene-d4	11.65	152	166057	49.60	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.20%

Target Compounds


Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl chloride	1.32	62	29290	10.852	ug/L	95
12) 1,1-Dichloroethene	2.13	96	8142	3.697	ug/L #	1
17) trans-1,2-Dichloroethene	2.78	96	16867	6.692	ug/L	97
19) 1,1-Dichloroethane	3.22	63	3955	0.896	ug/L #	92
20) cis-1,2-Dichloroethene	3.94	96	746033	278.522	ug/L	100
30) 1,1,1-Trichloroethane	4.64	97	7509	1.751	ug/L #	68
34) Trichloroethene	5.94	95	5183628	1759.088	ug/L	97
46) Tetrachloroethene	8.00	164	2326556	973.188	ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092920\
Data File : VV018550.D
Acq On : 30 Sep 2020 04:54
Operator : SY/MD
Sample : L4165-08
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AA8

Quant Time: Sep 30 06:09:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 30 04:50:47 2020
Response via : Initial Calibration





#5

Vinyl chloride

Concen: 10.852 ug/L

RT: 1.32 min Scan# 71

Delta R.T. 0.00 min

Lab File: VV018550.D

Acq: 30 Sep 2020 04:54

Instrument :

MSVOA_V

ClientSampleId :





C0AA8

Tot Ion: 62 Resp: 29290

Ion Ratio Lower Upper

62 100

64 35.5 22.9 42.5



#12

1,1-Dichloroethene

Concen: 3.697 ug/L

RT: 2.13 min Scan# 324

Delta R.T. 0.00 min

Lab File: VV018550.D

Acq: 30 Sep 2020 04:54




Tgt Ion: 96 Resp: 8142


Ion Ratio Lower Upper

96 100

61 270.4 122.8 228.1#

63 1160.2 91.9 170.7#





#17

trans-1,2-Dichloroethene

Concen: 6.692 ug/L

RT: 2.78 min Scan# 526

Delta R.T. 0.00 min

Lab File: VV018550.D

Acq: 30 Sep 2020 04:54

Instrument :

MSVOA_V

ClientSampleId :

C0AA8


Tot Ion: 96 Resp: 16867

Ion Ratio Lower Upper

96 100

61 133.4 95.0 176.4




98 66.7 44.6 82.8



Abundance Ion 96.00 (95.70 to 96.70): VV0

Ion 61.00 (60.70 to 61.70): VV0

Ion 98.00 (97.70 to 98.70): VV0



#19

1,1-Dichloroethane

Concen: 0.896 ug/L

RT: 3.22 min Scan# 662

Delta R.T. 0.01 min

Lab File: VV018550.D

Acq: 30 Sep 2020 04:54


Tgt Ion: 63 Resp: 3955

Ion Ratio Lower Upper

63 100

65 34.2 22.3 41.3



83 19.9 9.3 17.3#




Abundance Ion 63.00 (62.70 to 63.70): VV0

Ion 65.00 (64.70 to 65.70): VV0

Ion 83.00 (82.70 to 83.70): VV0









#20

cis-1,2-Dichloroethene
Concen: 278.522 ug/L
RT: 3.94 min Scan# 886
Delta R.T. 0.00 min
Lab File: VV018550.D
Acq: 30 Sep 2020 04:54

Instrument :
MSVOA_V
ClientSampled :
C0AA8




Tot Ion: 96 Resp: 746033
Ion Ratio Lower Upper
96 100
61 124.0 86.8 161.2
68 0.0 0.0 0.0




#30

1,1,1-Trichloroethane
Concen: 1.751 ug/L
RT: 4.64 min Scan# 1104
Delta R.T. 0.01 min
Lab File: VV018550.D
Acq: 30 Sep 2020 04:54

Tgt Ion: 97 Resp: 7509
Ion Ratio Lower Upper
97 100
99 32.6 51.4 77.0#
61 29.6 33.9 50.9#





#34

Trichloroethene

Concen: 1759.088 ug/L

RT: 5.94 min Scan# 1508

Delta R.T. 0.00 min

Lab File: VV018550.D

Acq: 30 Sep 2020 04:54

Instrument :

MSVOA_V

ClientSampleId :

C0AA8

Tot Ion: 95 Resp: 5183628


Ion Ratio Lower Upper

95 100

97 65.6 45.9 85.3

132 98.1 70.8 131.4

130 101.7 75.3 139.8






Abundance Ion 95.00 (94.70 to 95.70): VV0

Ion 97.00 (96.70 to 97.70): VV0

Ion 132.00 (131.70 to 132.70): V

Ion 130.00 (129.70 to 130.70): V



#46

Tetrachloroethene

Concen: 973.188 ug/L

RT: 8.00 min Scan# 2149

Delta R.T. 0.00 min

Lab File: VV018550.D

Acq: 30 Sep 2020 04:54

Tgt Ion: 164 Resp: 2326556


Ion Ratio Lower Upper

164 100

129 91.1 62.8 116.6

131 89.1 60.8 112.8

166 128.4 88.9 165.1



Abundance Ion 164.00 (163.70 to 164.70): V

Ion 129.00 (128.70 to 129.70): V

Ion 131.00 (130.70 to 131.70): V

Ion 166.00 (165.70 to 166.70): V

