

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092921\
 Data File : VV022375.D
 Acq On : 29 Sep 2021 14:08
 Operator : SY/MD
 Sample : M3978-20
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F4A64

Quant Time: Sep 30 05:28:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 30 05:14:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	251073	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	237060	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	106621	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	89524	42.015	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.040%
7) Chloroethane-d5	1.561	69	76425	45.348	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.700%
11) 1,1-Dichloroethene-d2	2.105	63	124786	33.684	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.360%
21) 2-Butanone-d5	3.873	46	144371	96.008	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.010%
24) Chloroform-d	4.349	84	174530	45.362	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.720%
26) 1,2-Dichloroethane-d4	5.034	65	109978	46.612	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.220%
32) Benzene-d6	5.050	84	358582	48.082	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.160%
36) 1,2-Dichloropropane-d6	6.069	67	116369	48.774	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.540%
41) Toluene-d8	7.317	98	313249	45.891	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.780%
43) trans-1,3-Dichloroprop...	7.622	79	48545	44.559	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.120%
47) 2-Hexanone-d5	8.088	63	89715	92.541	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.540%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	159551	46.069	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.140%
66) 1,2-Dichlorobenzene-d4	11.625	152	124315	52.435	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.860%
Target Compounds						
13) Acetone	2.159	43	26497	22.451	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092921\
Data File : VV022375.D
Acq On : 29 Sep 2021 14:08
Operator : SY/MD
Sample : M3978-20
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F4A64

Quant Time: Sep 30 05:28:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092721WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 30 05:14:53 2021
Response via : Initial Calibration

