

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092925\
 Data File : VV039222.D
 Acq On : 29 Sep 2025 11:38
 Operator : SY/MD
 Sample : VV0929MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV0929MBL01

Quant Time: Sep 30 01:22:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.597	168	503598	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	855660	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	847306	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	430776	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.941	65	373920	51.302	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	102.600%
35) Dibromofluoromethane	4.500	113	358071	52.054	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	104.100%
50) Toluene-d8	7.236	98	896223	44.361	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	88.720%#
62) 4-Bromofluorobenzene	10.001	95	370293	44.522	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	89.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092925\
Data File : VV039222.D
Acq On : 29 Sep 2025 11:38
Operator : SY/MD
Sample : VV0929MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV0929MBL01

Quant Time: Sep 30 01:22:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

