

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092925\
 Data File : VV039223.D
 Acq On : 29 Sep 2025 12:00
 Operator : SY/MD
 Sample : VV0929WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV0929WBL01

Quant Time: Sep 30 01:22:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	463653	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	797576	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.777	117	809645	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	403956	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	367686	54.793	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	109.580%
35) Dibromofluoromethane	4.503	113	349442	54.499	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	109.000%
50) Toluene-d8	7.240	98	860432	45.691	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	91.380%
62) 4-Bromofluorobenzene	10.002	95	353180	45.557	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	91.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092925\
Data File : VV039223.D
Acq On : 29 Sep 2025 12:00
Operator : SY/MD
Sample : VV0929WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV0929WBL01

Quant Time: Sep 30 01:22:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

