

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092925\
 Data File : VV039226.D
 Acq On : 29 Sep 2025 13:08
 Operator : SY/MD
 Sample : Q3195-15DL 5X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW-31S-091825DL

Quant Time: Sep 30 01:24:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	520194	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	885532	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	868036	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	445110	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	388329	51.579	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	= 103.160%	
35) Dibromofluoromethane	4.503	113	372446	52.317	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	= 104.640%	
50) Toluene-d8	7.239	98	941881	45.048	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	= 90.100%	
62) 4-Bromofluorobenzene	10.001	95	384328	44.651	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	= 89.300%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.082	101	357606	43.863	ug/l	96
44) Trichloroethene	5.841	130	40560	5.015	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092925\
Data File : VV039226.D
Acq On : 29 Sep 2025 13:08
Operator : SY/MD
Sample : Q3195-15DL 5X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW-31S-091825DL

Quant Time: Sep 30 01:24:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

