

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092925\
 Data File : VV039235.D
 Acq On : 29 Sep 2025 16:30
 Operator : SY/MD
 Sample : Q3231-12
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 WC-3

Quant Time: Sep 30 01:28:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 24 08:08:08 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	431521	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.535	114	765348	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	929145	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	397561	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.944	65	368061	58.933	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	117.860%
35) Dibromofluoromethane	4.503	113	350916	57.034	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	114.060%
50) Toluene-d8	7.236	98	976352	54.029	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	108.060%
62) 4-Bromofluorobenzene	10.001	95	356206	47.882	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	95.760%
Target Compounds						
16) Acetone	2.124	43	134753	29.212	ug/l	99
18) Methyl Acetate	2.388	43	20014	2.163	ug/l	91
20) Methylene Chloride	2.455	84	53966	6.142	ug/l	87
43) Isopropyl Acetate	5.211	43	29585	1.922	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092925\
Data File : VV039235.D
Acq On : 29 Sep 2025 16:30
Operator : SY/MD
Sample : Q3231-12
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
WC-3

Quant Time: Sep 30 01:28:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V092225W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 24 08:08:08 2025
Response via : Initial Calibration

