

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV093020\
 Data File : VV018566.D
 Acq On : 30 Sep 2020 13:54
 Operator : SY/MD
 Sample : L4149-07
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Oct 01 08:27:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 01 07:39:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	316749	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	308547	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	157982	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	80755	43.47	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.94%
7) Chloroethane-d5	1.58	69	77880	50.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.40%
11) 1,1-Dichloroethene-d2	2.12	63	138176	36.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.02%
21) 2-Butanone-d5	3.92	46	139077	100.37	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.37%
24) Chloroform-d	4.38	84	187745	46.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.48%
26) 1,2-Dichloroethane-d4	5.06	65	128500	50.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.78%
32) Benzene-d6	5.08	84	370123	47.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.72%
36) 1,2-Dichloropropane-d6	6.09	67	115155	48.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.52%
41) Toluene-d8	7.34	98	342917	46.64	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.28%
43) trans-1,3-Dichloropropene-	7.64	79	59977	47.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.04%
47) 2-Hexanone-d5	8.11	63	79203	88.30	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.30%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	142224	45.43	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.86%
64) 1,2-Dichlorobenzene-d4	11.65	152	154902	50.98	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.96%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV093020\
Data File : VV018566.D
Acq On : 30 Sep 2020 13:54
Operator : SY/MD
Sample : L4149-07
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Oct 01 08:27:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 01 07:39:17 2020
Response via : Initial Calibration

