

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV093020\
 Data File : VV018556.D
 Acq On : 30 Sep 2020 09:58
 Operator : SY/MD
 Sample : VV0930MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK88

Quant Time: Oct 01 07:52:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 01 07:39:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	364560	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	360200	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	183859	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	99974	46.76	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.52%
7) Chloroethane-d5	1.58	69	97415	55.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	110.20%
11) 1,1-Dichloroethene-d2	2.12	63	167126	38.36	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.72%
21) 2-Butanone-d5	3.92	46	160749	100.79	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.79%
24) Chloroform-d	4.37	84	218565	47.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.54%
26) 1,2-Dichloroethane-d4	5.05	65	147629	50.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.60%
32) Benzene-d6	5.07	84	440351	48.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.52%
36) 1,2-Dichloropropane-d6	6.09	67	136800	49.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.24%
41) Toluene-d8	7.33	98	407126	47.43	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.86%
43) trans-1,3-Dichloropropene-	7.64	79	70254	47.18	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.36%
47) 2-Hexanone-d5	8.11	63	92485	88.33	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.33%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	166447	45.54	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.08%
64) 1,2-Dichlorobenzene-d4	11.65	152	180299	50.99	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.98%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV093020\
Data File : VV018556.D
Acq On : 30 Sep 2020 09:58
Operator : SY/MD
Sample : VV0930MBL01
Misc : 5.00µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK88

Quant Time: Oct 01 07:52:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 01 07:39:17 2020
Response via : Initial Calibration

