

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV093020\
 Data File : VV018557.D
 Acq On : 30 Sep 2020 10:22
 Operator : SY/MD
 Sample : VV0930WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK89

Quant Time: Oct 01 07:55:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 01 07:39:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	356261	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	346167	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	175342	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	97338	46.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.18%
7) Chloroethane-d5	1.58	69	92435	53.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	107.00%
11) 1,1-Dichloroethene-d2	2.12	63	163943	38.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.02%
21) 2-Butanone-d5	3.92	46	156768	100.59	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.59%
24) Chloroform-d	4.38	84	211907	46.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.80%
26) 1,2-Dichloroethane-d4	5.06	65	146197	50.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.94%
32) Benzene-d6	5.07	84	426860	48.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.36%
36) 1,2-Dichloropropane-d6	6.09	67	134839	50.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.78%
41) Toluene-d8	7.34	98	393293	47.68	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.36%
43) trans-1,3-Dichloropropene-	7.64	79	66853	46.71	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.42%
47) 2-Hexanone-d5	8.11	63	90018	89.45	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.45%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	161918	46.10	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.20%
64) 1,2-Dichlorobenzene-d4	11.65	152	174328	51.70	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV093020\
Data File : VV018557.D
Acq On : 30 Sep 2020 10:22
Operator : SY/MD
Sample : VV0930WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK89

Quant Time: Oct 01 07:55:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 01 07:39:17 2020
Response via : Initial Calibration

