

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW093021\
 Data File : VW022457.D
 Acq On : 01 Oct 2021 01:42
 Operator : SY/MD
 Sample : M3978-21
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Oct 01 04:47:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 01 04:27:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	217413	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	210635	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	91414	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	86294	46.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.54%
7) Chloroethane-d5	1.561	69	70260	48.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.28%
11) 1,1-Dichloroethene-d2	2.105	63	115303	35.94	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.88%
21) 2-Butanone-d5	3.876	46	122989	94.45	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.45%
24) Chloroform-d	4.349	84	144144	43.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.52%
26) 1,2-Dichloroethane-d4	5.031	65	92150	45.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.20%
32) Benzene-d6	5.050	84	305282	46.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.14%
36) 1,2-Dichloropropane-d6	6.069	67	96895	45.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.42%
41) Toluene-d8	7.317	98	269389	44.42	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.84%
43) trans-1,3-Dichloroprop...	7.622	79	38406	39.67	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.34%
47) 2-Hexanone-d5	8.088	63	74832	86.87	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.87%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	128723	41.83	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.66%
66) 1,2-Dichlorobenzene-d4	11.625	152	102716	50.53	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.06%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV093021\
Data File : VV022457.D
Acq On : 01 Oct 2021 01:42
Operator : SY/MD
Sample : M3978-21
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Oct 01 04:47:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092721WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 01 04:27:45 2021
Response via : Initial Calibration

