

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV093021\
 Data File : VV022466.D
 Acq On : 01 Oct 2021 05:18
 Operator : SY/MD
 Sample : M3997-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F4A47

Quant Time: Oct 01 06:28:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 01 04:27:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	226429	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	218248	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	96050	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	81988	42.67	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	85.34%		
7) Chloroethane-d5	1.555	69	69154	45.50	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	91.00%		
11) 1,1-Dichloroethene-d2	2.101	63	113498	33.97	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	67.94%		
21) 2-Butanone-d5	3.873	46	106315	78.40	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	78.40%		
24) Chloroform-d	4.346	84	145997	42.08	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	84.16%		
26) 1,2-Dichloroethane-d4	5.031	65	90075	42.33	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	84.66%		
32) Benzene-d6	5.050	84	299589	43.63	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	87.26%		
36) 1,2-Dichloropropane-d6	6.069	67	96177	43.79	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	87.58%		
41) Toluene-d8	7.317	98	261696	41.64	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	83.28%		
43) trans-1,3-Dichloroprop...	7.622	79	35348	35.24	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	70.48%		
47) 2-Hexanone-d5	8.088	63	63586	71.24	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	71.24%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	124526	39.05	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	78.10%		
66) 1,2-Dichlorobenzene-d4	11.625	152	103952	48.67	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	97.34%		
Target Compounds						
					Qvalue	
17) trans-1,2-Dichloroethene	2.757	96	29818	17.36	ug/L	95
18) Methyl tert-butyl Ether	2.767	73	22200	4.35	ug/L	98
20) cis-1,2-Dichloroethene	3.915	96	14910	8.15	ug/L	95
46) Tetrachloroethene	7.976	164	61567	43.66	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV093021\
Data File : VV022466.D
Acq On : 01 Oct 2021 05:18
Operator : SY/MD
Sample : M3997-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F4A47

Quant Time: Oct 01 06:28:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092721WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 01 04:27:45 2021
Response via : Initial Calibration

