

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW093021\
 Data File : VW022456.D
 Acq On : 01 Oct 2021 01:18
 Operator : SY/MD
 Sample : M3979-21
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleID :
 VHBLK001

Quant Time: Oct 01 04:47:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 01 04:27:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	222438	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	213865	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	93386	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	88897	47.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.18%
7) Chloroethane-d5	1.561	69	73395	49.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.32%
11) 1,1-Dichloroethene-d2	2.105	63	119419	36.39	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.78%
21) 2-Butanone-d5	3.876	46	121229	91.00	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.00%
24) Chloroform-d	4.349	84	149549	43.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.74%
26) 1,2-Dichloroethane-d4	5.034	65	94434	45.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.36%
32) Benzene-d6	5.050	84	314893	46.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.60%
36) 1,2-Dichloropropane-d6	6.069	67	99791	46.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.72%
41) Toluene-d8	7.317	98	274249	44.53	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.06%
43) trans-1,3-Dichloroprop...	7.625	79	37284	37.93	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.86%
47) 2-Hexanone-d5	8.092	63	73474	84.01	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.01%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	129338	41.40	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.80%
66) 1,2-Dichlorobenzene-d4	11.625	152	104132	50.15	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.30%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV093021\
Data File : VV022456.D
Acq On : 01 Oct 2021 01:18
Operator : SY/MD
Sample : M3979-21
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Oct 01 04:47:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092721WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 01 04:27:45 2021
Response via : Initial Calibration

