

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100118\
 Data File : VV007995.D
 Acq On : 01 Oct 2018 15:22
 Operator : SY/MD
 Sample : J5063-23
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Oct 02 03:49:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 02 02:56:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	85299	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	80500	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	39797	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	24856	5.43	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.60%
7) Chloroethane-d5	1.59	69	21060	5.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.60%
11) 1,1-Dichloroethene-d2	2.14	63	39073	3.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.60%
20) 2-Butanone-d5	3.96	46	66648	48.46	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.92%
24) Chloroform-d	4.41	84	54087	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.80%
26) 1,2-Dichloroethane-d4	5.09	65	29002	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.20%
32) Benzene-d6	5.10	84	105536	5.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.80%
36) 1,2-Dichloropropane-d6	6.13	67	32586	5.35	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.00%
41) Toluene-d8	7.36	98	101243	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.40%
43) trans-1,3-Dichloropropene-	7.67	79	11426	4.87	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.40%
46) 2-Hexanone-d5	8.14	63	48055	46.41	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.82%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	23609	5.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	41491	5.39	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100118\
Data File : VV007995.D
Acq On : 01 Oct 2018 15:22
Operator : SY/MD
Sample : J5063-23
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Oct 02 03:49:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 02 02:56:09 2018
Response via : Initial Calibration

