

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100120\
 Data File : VV018583.D
 Acq On : 01 Oct 2020 11:44
 Operator : SY/MD
 Sample : VV1001WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK91

Quant Time: Oct 01 23:47:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 01 23:37:24 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	350681	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	328351	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	170134	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	120152	58.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	116.84%
7) Chloroethane-d5	1.58	69	107231	63.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	126.12%
11) 1,1-Dichloroethene-d2	2.12	63	182811	43.62	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.24%
21) 2-Butanone-d5	3.92	46	146823	95.71	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.71%
24) Chloroform-d	4.38	84	213797	48.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.14%
26) 1,2-Dichloroethane-d4	5.06	65	150303	53.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.48%
32) Benzene-d6	5.08	84	465928	56.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.04%
36) 1,2-Dichloropropane-d6	6.09	67	139779	55.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	111.24%
41) Toluene-d8	7.34	98	406035	51.89	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.78%
43) trans-1,3-Dichloropropene-	7.64	79	68204	50.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	100.48%
47) 2-Hexanone-d5	8.11	63	84828	88.87	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.87%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	156335	46.92	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.84%
64) 1,2-Dichlorobenzene-d4	11.65	152	175556	53.66	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.32%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100120\
Data File : VV018583.D
Acq On : 01 Oct 2020 11:44
Operator : SY/MD
Sample : VV1001WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK91

Quant Time: Oct 01 23:47:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 01 23:37:24 2020
Response via : Initial Calibration

