

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100120\
 Data File : VV018606.D
 Acq On : 01 Oct 2020 22:21
 Operator : SY/MD
 Sample : L4237-08
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0AC3

Quant Time: Oct 02 05:07:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 02 04:18:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	289814	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	282831	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	140431	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	86424	50.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.70%
7) Chloroethane-d5	1.58	69	81537	58.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	116.04%
11) 1,1-Dichloroethene-d2	2.12	63	144262	41.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.32%
21) 2-Butanone-d5	3.92	46	113020	89.14	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.14%
24) Chloroform-d	4.38	84	169915	46.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.46%
26) 1,2-Dichloroethane-d4	5.06	65	117513	50.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.72%
32) Benzene-d6	5.07	84	337073	47.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.10%
36) 1,2-Dichloropropane-d6	6.09	67	101033	46.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.34%
41) Toluene-d8	7.34	98	302582	44.90	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.80%
43) trans-1,3-Dichloropropene-	7.64	79	46924	40.13	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.26%
47) 2-Hexanone-d5	8.11	63	60881	74.05	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	74.05%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	128048	44.62	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.24%
64) 1,2-Dichlorobenzene-d4	11.65	152	131598	48.73	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.46%

Target Compounds					Ovalue
34) Trichloroethene	5.95	95	2234	0.926 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100120\
Data File : VV018606.D
Acq On : 01 Oct 2020 22:21
Operator : SY/MD
Sample : L4237-08
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0AC3

Quant Time: Oct 02 05:07:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 02 04:18:06 2020
Response via : Initial Calibration

