

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100121\
 Data File : VV022488.D
 Acq On : 02 Oct 2021 01:40
 Operator : SY/MD
 Sample : M4024-08
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AJ5

Quant Time: Oct 02 03:33:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 02 01:27:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	230428	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	218119	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	100207	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	65479	45.947	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.900%
7) Chloroethane-d5	1.568	69	55462	49.510	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.020%
11) 1,1-Dichloroethene-d2	2.108	63	94212	33.528	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.060%
21) 2-Butanone-d5	3.896	46	61470	78.361	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	78.360%
24) Chloroform-d	4.349	84	120804	46.397	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.800%
26) 1,2-Dichloroethane-d4	5.034	65	74747	49.488	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.980%
32) Benzene-d6	5.050	84	245768	48.637	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.280%
36) 1,2-Dichloropropane-d6	6.072	67	75653	48.633	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.260%
41) Toluene-d8	7.317	98	217340	46.890	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.780%
43) trans-1,3-Dichloroprop...	7.625	79	31756	43.147	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.300%
47) 2-Hexanone-d5	8.092	63	61780	95.871	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.870%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	101516	46.931	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.860%
66) 1,2-Dichlorobenzene-d4	11.625	152	84076	51.450	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.900%
Target Compounds						Qvalue
61) 1,2,3-Trichloropropane	11.249	75	11209	5.882	ug/L	# 62


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100121\
Data File : VV022488.D
Acq On : 02 Oct 2021 01:40
Operator : SY/MD
Sample : M4024-08
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AJ5


Quant Time: Oct 02 03:33:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100121WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 02 01:27:14 2021
Response via : Initial Calibration





#61
 1,2,3-Trichloropropane
 Concen: 5.882 ug/L
 RT: 11.249 min Scan# 3175
 Delta R.T. 0.974 min
 Lab File: VV022488.D
 Acq: 02 Oct 2021 01:40

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AJ5



Tgt Ion: 75	Resp: 11209
Ion Ratio	Lower Upper
75	100
77	33.4 26.0 39.0
110	1.2 34.3 51.5#

