

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100121\
 Data File : VV022483.D
 Acq On : 01 Oct 2021 23:41
 Operator : SY/MD
 Sample : M4024-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AH1

Quant Time: Oct 02 01:38:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 02 01:27:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	240264	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	223592	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	103739	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	70345	47.341	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.680%
7) Chloroethane-d5	1.565	69	57701	49.400	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.800%
11) 1,1-Dichloroethene-d2	2.105	63	103839	35.441	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.880%
21) 2-Butanone-d5	3.892	46	70267	85.908	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.910%
24) Chloroform-d	4.349	84	121951	44.920	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.840%
26) 1,2-Dichloroethane-d4	5.034	65	74861	47.535	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.060%
32) Benzene-d6	5.050	84	255865	49.396	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.800%
36) 1,2-Dichloropropane-d6	6.072	67	78257	49.076	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.160%
41) Toluene-d8	7.317	98	228265	48.041	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.080%
43) trans-1,3-Dichloroprop...	7.622	79	32416	42.966	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.940%
47) 2-Hexanone-d5	8.092	63	53818	81.471	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.470%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	99453	44.851	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.700%
66) 1,2-Dichlorobenzene-d4	11.625	152	86510	51.137	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.280%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.111	96	1709	1.018	ug/L	# 1
20) cis-1,2-Dichloroethene	3.912	96	15749	8.024	ug/L	96
34) Trichloroethene	5.928	95	1921	1.066	ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100121\
Data File : VV022483.D
Acq On : 01 Oct 2021 23:41
Operator : SY/MD
Sample : M4024-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AH1

Quant Time: Oct 02 01:38:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100121WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 02 01:27:14 2021
Response via : Initial Calibration

