

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100218\
 Data File : VV008024.D
 Acq On : 02 Oct 2018 11:28
 Operator : SY/MD
 Sample : VV1002WBL01
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK75

Quant Time: Oct 03 05:41:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 03 05:34:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	84865	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	81040	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	39271	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	25777	5.66	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	113.20%
7) Chloroethane-d5	1.59	69	22535	5.73	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	114.60%
11) 1,1-Dichloroethene-d2	2.13	63	40625	3.95	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.00%
20) 2-Butanone-d5	3.98	46	72270	52.81	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.62%
24) Chloroform-d	4.41	84	57214	5.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.40%
26) 1,2-Dichloroethane-d4	5.09	65	30059	5.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.40%
32) Benzene-d6	5.10	84	112773	5.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	116.60%
36) 1,2-Dichloropropane-d6	6.12	67	35163	5.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	114.60%
41) Toluene-d8	7.36	98	104789	5.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.40%
43) trans-1,3-Dichloropropene-	7.67	79	12044	5.10	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	102.00%
46) 2-Hexanone-d5	8.14	63	53167	51.00	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.00%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	23596	4.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	42609	5.61	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100218\
Data File : VV008024.D
Acq On : 02 Oct 2018 11:28
Operator : SY/MD
Sample : VV1002WBL01
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK75

Quant Time: Oct 03 05:41:52 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 03 05:34:53 2018
Response via : Initial Calibration

