

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100218\
 Data File : VV008040.D
 Acq On : 02 Oct 2018 20:45
 Operator : SY/MD
 Sample : J5185-10
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JLC88

Quant Time: Oct 03 07:03:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 03 05:34:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	85397	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	81720	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	39258	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	22145	4.84	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.80%
7) Chloroethane-d5	1.59	69	20053	5.07	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.40%
11) 1,1-Dichloroethene-d2	2.13	63	34906	3.38	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.60%
20) 2-Butanone-d5	3.96	46	67336	48.90	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.80%
24) Chloroform-d	4.41	84	52010	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.80%
26) 1,2-Dichloroethane-d4	5.09	65	28463	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.20%
32) Benzene-d6	5.10	84	99286	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.80%
36) 1,2-Dichloropropane-d6	6.12	67	32702	5.29	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.80%
41) Toluene-d8	7.36	98	92250	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.80%
43) trans-1,3-Dichloropropene-	7.67	79	10426	4.38	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.60%
46) 2-Hexanone-d5	8.14	63	44779	42.60	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.20%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	21771	4.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	38805	5.11	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100218\
Data File : VV008040.D
Acq On : 02 Oct 2018 20:45
Operator : SY/MD
Sample : J5185-10
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JLC88

Quant Time: Oct 03 07:03:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 03 05:34:53 2018
Response via : Initial Calibration

