

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100218\
 Data File : VV008031.D
 Acq On : 02 Oct 2018 16:10
 Operator : SY/MD
 Sample : J5185-01
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JLC77

Quant Time: Oct 03 06:41:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 03 05:34:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	107309	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	102648	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	49623	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	21954	3.82	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.40%
7) Chloroethane-d5	1.59	69	20090	4.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.80%
11) 1,1-Dichloroethene-d2	2.13	63	35877	2.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	55.20%#
20) 2-Butanone-d5	3.96	46	61269	35.41	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	70.82%
24) Chloroform-d	4.41	84	50375	3.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.80%
26) 1,2-Dichloroethane-d4	5.09	65	26637	3.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.00%
32) Benzene-d6	5.10	84	96676	3.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.00%
36) 1,2-Dichloropropane-d6	6.12	67	30786	3.96	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	79.20%
41) Toluene-d8	7.36	98	93528	3.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.00%
43) trans-1,3-Dichloropropene-	7.67	79	10302	3.44	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	68.80%
46) 2-Hexanone-d5	8.14	63	43265	32.77	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	65.54%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	20916	3.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	69.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	37982	3.95	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	79.00%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100218\
Data File : VV008031.D
Acq On : 02 Oct 2018 16:10
Operator : SY/MD
Sample : J5185-01
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JLC77

Quant Time: Oct 03 06:41:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 03 05:34:53 2018
Response via : Initial Calibration

