

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100220\
 Data File : VV018617.D
 Acq On : 02 Oct 2020 11:50
 Operator : SY/MD
 Sample : VV1002MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK99

Quant Time: Oct 03 01:09:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 03 00:50:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	282554	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	281004	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	142997	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	85693	51.71	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	103.42%
7) Chloroethane-d5	1.58	69	79725	58.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	116.38%
11) 1,1-Dichloroethene-d2	2.12	63	144885	42.91	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.82%
21) 2-Butanone-d5	3.90	46	110524	89.41	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.41%
24) Chloroform-d	4.37	84	168452	47.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.02%
26) 1,2-Dichloroethane-d4	5.05	65	113576	49.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.86%
32) Benzene-d6	5.07	84	331222	46.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.06%
36) 1,2-Dichloropropane-d6	6.09	67	100674	46.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.62%
41) Toluene-d8	7.33	98	298993	44.65	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.30%
43) trans-1,3-Dichloropropene-	7.64	79	49862	42.92	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.84%
47) 2-Hexanone-d5	8.11	63	61414	75.18	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	75.18%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	119077	41.76	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.52%
64) 1,2-Dichlorobenzene-d4	11.65	152	132619	48.22	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.44%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100220\
Data File : VV018617.D
Acq On : 02 Oct 2020 11:50
Operator : SY/MD
Sample : VV1002MBL01
Misc : 5.00µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK99

Quant Time: Oct 03 01:09:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 03 00:50:30 2020
Response via : Initial Calibration

