

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100220\
 Data File : VV018648.D
 Acq On : 03 Oct 2020 00:19
 Operator : SY/MD
 Sample : L4254-02 100X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3Q6

Quant Time: Oct 03 04:36:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 03 02:08:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	299861	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	296745	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	144537	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	78508	44.64	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.28%
7) Chloroethane-d5	1.58	69	68639	47.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.40%
11) 1,1-Dichloroethene-d2	2.12	63	126528	35.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.62%
21) 2-Butanone-d5	3.90	46	107991	82.32	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	82.32%
24) Chloroform-d	4.38	84	163372	42.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.92%
26) 1,2-Dichloroethane-d4	5.06	65	112279	46.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.02%
32) Benzene-d6	5.07	84	313624	41.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.44%
36) 1,2-Dichloropropane-d6	6.09	67	96212	42.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	84.72%
41) Toluene-d8	7.34	98	282933	40.01	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.02%
43) trans-1,3-Dichloropropene-	7.64	79	47041	38.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.68%
47) 2-Hexanone-d5	8.11	63	60697	70.36	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	70.36%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	117051	38.88	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	77.76%
64) 1,2-Dichlorobenzene-d4	11.65	152	128037	46.06	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.12%

Target Compounds					Ovalue
46) Tetrachloroethene	8.00	164	22962	11.183 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100220\
Data File : VV018648.D
Acq On : 03 Oct 2020 00:19
Operator : SY/MD
Sample : L4254-02 100X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3Q6

Quant Time: Oct 03 04:36:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 03 02:08:43 2020
Response via : Initial Calibration

