

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100220\
 Data File : VV018618.D
 Acq On : 02 Oct 2020 12:16
 Operator : SY/MD
 Sample : VV1002WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK61

Quant Time: Oct 03 01:12:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 03 00:50:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	277579	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	277155	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	138872	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	81209	49.88	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	99.76%
7) Chloroethane-d5	1.58	69	80003	59.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	118.88%
11) 1,1-Dichloroethene-d2	2.12	63	142925	43.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.18%
21) 2-Butanone-d5	3.91	46	106599	87.79	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	87.79%
24) Chloroform-d	4.37	84	167623	47.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.24%
26) 1,2-Dichloroethane-d4	5.05	65	113679	50.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.74%
32) Benzene-d6	5.07	84	324772	46.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.52%
36) 1,2-Dichloropropane-d6	6.09	67	100792	47.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.04%
41) Toluene-d8	7.34	98	292294	44.26	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.52%
43) trans-1,3-Dichloropropene-	7.64	79	50488	44.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.12%
47) 2-Hexanone-d5	8.11	63	57300	71.12	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	71.12%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	111276	39.57	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	79.14%
64) 1,2-Dichlorobenzene-d4	11.65	152	132387	49.57	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.14%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100220\
Data File : VV018618.D
Acq On : 02 Oct 2020 12:16
Operator : SY/MD
Sample : VV1002WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK61

Quant Time: Oct 03 01:12:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 03 00:50:30 2020
Response via : Initial Calibration

