

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100223\
 Data File : VV032202.D
 Acq On : 02 Oct 2023 16:36
 Operator : SY/MD
 Sample : VIBLK233
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK233

Quant Time: Oct 03 02:01:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 03 01:57:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	187634	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	183970	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	90276	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	39024	33.607	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	67.220%		
7) Chloroethane-d5	1.529	69	40024	39.962	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	79.920%		
11) 1,1-Dichloroethene-d2	2.056	65	17928	35.358	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	70.720%		
21) 2-Butanone-d5	3.786	46	102709	98.372	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	98.370%		
24) Chloroform-d	4.249	84	104618	43.054	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	86.100%		
26) 1,2-Dichloroethane-d4	4.941	65	69076	47.247	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.500%		
32) Benzene-d6	4.960	84	198916	41.130	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	82.260%		
36) 1,2-Dichloropropane-d6	5.989	67	73051	45.264	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	90.520%		
41) Toluene-d8	7.246	98	166431	37.722	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	75.440%#		
43) trans-1,3-Dichloroprop...	7.555	79	30225	39.368	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	78.740%		
47) 2-Hexanone-d5	8.027	63	58853	84.446	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	84.450%		
56) 1,1,2,2-Tetrachloroeth...	10.156	84	103665	45.162	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	90.320%		
66) 1,2-Dichlorobenzene-d4	11.564	152	72392	44.219	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	88.440%		
Target Compounds					Qvalue	
13) Acetone	2.114	43	5461	5.591	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100223\
Data File : VV032202.D
Acq On : 02 Oct 2023 16:36
Operator : SY/MD
Sample : VIBLK233
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK233

Quant Time: Oct 03 02:01:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092523WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 03 01:57:58 2023
Response via : Initial Calibration

