

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100318\
 Data File : VV008051.D
 Acq On : 03 Oct 2018 15:44
 Operator : SY/MD
 Sample : J5185-13
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBKL01

Quant Time: Oct 04 08:26:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Oct 04 07:54:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	81267	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	76372	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	38356	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	20950	4.81	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.20%
7) Chloroethane-d5	1.59	69	20250	5.38	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.60%
11) 1,1-Dichloroethene-d2	2.14	63	35388	3.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.00%
20) 2-Butanone-d5	3.96	46	70907	54.11	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.22%
24) Chloroform-d	4.41	84	53740	5.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.40%
26) 1,2-Dichloroethane-d4	5.09	65	28422	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.20%
32) Benzene-d6	5.10	84	102220	5.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.20%
36) 1,2-Dichloropropane-d6	6.13	67	33270	5.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	115.20%
41) Toluene-d8	7.36	98	94155	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.40%
43) trans-1,3-Dichloropropene-	7.67	79	10935	4.91	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.20%
46) 2-Hexanone-d5	8.14	63	50203	51.10	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.20%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	23361	5.24	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	39377	5.30	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100318\
Data File : VV008051.D
Acq On : 03 Oct 2018 15:44
Operator : SY/MD
Sample : J5185-13
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBKL01

Quant Time: Oct 04 08:26:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR092718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Oct 04 07:54:45 2018
Response via : Initial Calibration

