

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100320\
 Data File : VV018660.D
 Acq On : 03 Oct 2020 17:20
 Operator : SY/MD
 Sample : L4203-06
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Oct 05 05:32:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Oct 05 04:50:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	307110	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	307419	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	147795	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	93414	51.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	103.72%
7) Chloroethane-d5	1.58	69	82108	55.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	110.26%
11) 1,1-Dichloroethene-d2	2.12	63	145539	39.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.32%
21) 2-Butanone-d5	3.92	46	111769	83.19	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.19%
24) Chloroform-d	4.38	84	176324	45.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.54%
26) 1,2-Dichloroethane-d4	5.06	65	119828	48.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.92%
32) Benzene-d6	5.07	84	336444	43.21	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.42%
36) 1,2-Dichloropropane-d6	6.09	67	105207	44.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.44%
41) Toluene-d8	7.34	98	304668	41.59	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.18%
43) trans-1,3-Dichloropropene-	7.64	79	50475	39.71	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.42%
47) 2-Hexanone-d5	8.11	63	59000	66.02	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	66.02%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	115673	37.08	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	74.16%
64) 1,2-Dichlorobenzene-d4	11.65	152	135331	47.61	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.22%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100320\
Data File : VV018660.D
Acq On : 03 Oct 2020 17:20
Operator : SY/MD
Sample : L4203-06
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Oct 05 05:32:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Oct 05 04:50:26 2020
Response via : Initial Calibration

