

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV100320\
 Data File : VV018662.D
 Acq On : 03 Oct 2020 18:07
 Operator : SY/MD
 Sample : L4237-18
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Oct 05 05:39:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Oct 05 04:50:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	313458	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	309531	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	148045	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	81682	44.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.86%
7) Chloroethane-d5	1.58	69	76536	50.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.70%
11) 1,1-Dichloroethene-d2	2.12	63	134939	36.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.04%
21) 2-Butanone-d5	3.92	46	108865	79.39	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	79.39%
24) Chloroform-d	4.38	84	173829	43.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.46%
26) 1,2-Dichloroethane-d4	5.06	65	117277	46.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.94%
32) Benzene-d6	5.08	84	333823	42.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.16%
36) 1,2-Dichloropropane-d6	6.09	67	104582	44.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.30%
41) Toluene-d8	7.34	98	299285	40.58	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.16%
43) trans-1,3-Dichloropropene-	7.64	79	50999	39.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.70%
47) 2-Hexanone-d5	8.11	63	58557	65.08	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	65.08%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	114979	36.61	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	73.22%
64) 1,2-Dichlorobenzene-d4	11.65	152	135919	47.74	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.48%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV100320\
Data File : VV018662.D
Acq On : 03 Oct 2020 18:07
Operator : SY/MD
Sample : L4237-18
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Oct 05 05:39:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM092920WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Oct 05 04:50:26 2020
Response via : Initial Calibration

